



Mobile Bath Lift System

Hygiene Equipment
Designed for People



Generic Specification text for a Mobile Bath Lift System

Specification for Mobile bath lift system (battery operated)

Generic text for Chair, Stretcher and Combi models

- Equipment shall be delivered 100 % assembled and ready for use after charging the batteries
- The product is height-adjustable for caregiver's ergonomic use and is specifically designed and manufactured for safe in-patient hospital use
- The lift shall offer a dignified and secure way for disabled patients to be lifted, transferred, seated and transported into and from the bathtub
- Foot operated lifting system for raising and lowering to assure hands free operation for caregiver
- Patients shall have the possibility to be lifted from the floor and beds with the mobile bath lift
- Compact base design to allow access under low bed frames and bathtub frames
- The lift is belt driven, runs gently, quietly and operates free from sudden movements
- Foot activated emergency stop
- Equipped with two brake castors and two straight steering castors for safe and easy patient transport
- Automated fall stop in case of components malfunction
- High capacity rechargeable batteries (min. 24 V, 7Ah) to be charged via main supply
- Magnetic battery charger connection for safety and ease of use
- Seating surfaces shall be made of a soft, durable material that is non water absorbent, easy to clean and disinfect and comfortable to sit and lay on
- Stainless steel lifting column
- Anti-corrosion treatment on all sheet metal parts

Specification text to add for chair lift


- Chair lift shall have a mirror-design for pick-up from either side of the bed
- A minimum of 2 safety belts to secure the patient

Specification text to add for stretcher lift

- Stretcher lift shall transfer both sitting and recumbent patients
- A minimum of 3 safety belts to secure the patient
- Side parts adjustable in variable height levels for patient comfort
- Stretcher lift shall have a mirror-design for pick-up from either side of the bed



TR EQUIPMENT



Mobile Bath Lift System

Hygiene Equipment
Designed for People



Specification text to add for combi lift


- Mobile lift needs to be easily converted from chair to stretcher lift and vice versa without tools
- Mobile lift shall transfer both sitting and recumbent patients
- A minimum of 3 safety belts to secure the patient
- Side parts adjustable in variable height levels for patient comfort
- Mobile lift shall have a mirror-design for pick-up from either side of the bed
- Wall bracket for chair and stretcher

Specification text to add for options/accessories if needed

- Both hand and foot operated control of height adjustment for caregiver convenience
- Scale for lift

Load specifications and operating height range

- Max. Patient Weight: 150 kg /330 lbs.
- Operating Height Range 200 - 965 mm / 8 - 38"



TR EQUIPMENT